



CE Certificate of Conformity

APPLICANT: ACCU-SCOPE, Inc., 73 Mall Drive, Commack, NY 11725

MANUFACTURER: ACCU-SCOPE, Inc., 73 Mall Drive, Commack, NY 11725

PRODUCT: Compound Microscope

MODEL: 3000-LED, 3001-LED, 3000-LED-SPH, 3001-LED-SPH, 3000-LED-40, 3001-LED-40, 3000-LED-10, 3001-LED-10
EXC-350, EXC-350-APO, EXC-350-PH, EXC-350-SPH, EXC-350-470, EXC-350-455, EXC-350-365, EXC-350-560

RATINGS: AC 100-240v, 50/60Hz

REPORT NO: EED45K000504

TEST STANDARDS: EN 61326-1: 2013
CISPR11: 2015
IEC 61000-3-2: 2014
IEC 61000-3-3: 2013

CERTIFICATE ISSUE DATE: August 19, 2019

DECLARATION ISSUED BY: Centre Testing International Group Co., LTD.
Hongwei Industrial Zone, Bao'an 70 District, Shenzhen, Guangdong, China

This Declaration is for the exclusive use of CTI's Client and is provided pursuant to the agreement between CTI and its Client. The observations and test results from this Declaration are relevant only to the sample tested. This Declaration by itself does not imply that the material, product, or service is or has ever been under a CTI certification program.

NOTE: This Declaration is part of the full test report(s) and should be read in conjunction with it.



Brian Taub
Executive Vice President
ACCU-SCOPE, Inc.

Legally binding signature of the issuer

